

[Home](#)[Online Library ACP](#)[Recent Final Revised Papers](#)[Volumes and Issues](#)[Special Issues](#)[Library Search](#)[Title and Author Search](#)[Online Library ACPD](#)[Alerts & RSS Feeds](#)[General Information](#)[Submission](#)[Review](#)[Production](#)[Subscription](#)[Comment on a Paper](#)[Volumes and Issues](#) [Contents of Issue 5](#)

Atmos. Chem. Phys., 8, 1153–1179, 2008

[www.atmos-chem-phys.net/8/1153/2008/](http://www.atmos-chem-phys.net/8/1153/2008/)

© Author(s) 2008. This work is distributed

under the Creative Commons Attribution 3.0 License.

## Calibration and measurement uncertainties of a continuous-flow cloud condensation nuclei counter (DMT-CCNC): CCN activation of ammonium sulfate and sodium chloride aerosol particles in theory and experiment

D. Rose<sup>1</sup>, S. S. Gunthe<sup>1</sup>, E. Mikhailov<sup>1,2</sup>, G. P. Frank<sup>1,\*</sup>, U. Dusek<sup>1,\*\*</sup>, M. O. Andreae<sup>1</sup>, and U. Pöschl<sup>1</sup><sup>1</sup>Max Planck Institute for Chemistry, Biogeochemistry Department, P.O. Box 3060, 55020 Mainz, Germany<sup>2</sup>Atmospheric Physics Department, Institute of Physics, St. Petersburg State University, 198904 St. Petersburg, Russia

\* now at: Lund University, Department of Physics, Lund, Sweden

\*\* now at: IMAU, Utrecht University, Utrecht, The Netherlands

**Abstract.** Experimental and theoretical uncertainties in the measurement of cloud condensation nuclei (CCN) with a continuous-flow thermal-gradient CCN counter from Droplet Measurement Technologies (DMT-CCNC) have been assessed by model calculations and calibration experiments with ammonium sulfate and sodium chloride aerosol particles in the diameter range of 20–220 nm. Experiments have been performed in the laboratory and during field measurement campaigns, covering a wide range of instrument operating conditions (650–1020 hPa pressure, 293–303 K inlet temperature, 4–34 K m<sup>-1</sup> temperature gradient, 0.5–1.0 L min<sup>-1</sup> flow rate). For each set of conditions, the effective water vapor supersaturation ( $S_{\text{eff}}$ , 0.05–1.4%) was determined from the measured CCN activation spectra (dry particle activation diameters) and Köhler model calculations. High measurement precision was achieved under stable laboratory conditions, where the relative standard deviations of  $S_{\text{eff}}$  were as low as  $\pm 1\%$ . During field measurements, however, the relative deviations increased to about  $\pm 5\%$ , which can be mostly attributed to variations of the CCNC column top temperature with ambient temperature. The observed dependence of  $S_{\text{eff}}$  on temperature, pressure, and flow rate was compared to the CCNC flow model of Lance et al. (2006). At high  $S_{\text{eff}}$  the relative deviations between flow model and experimental results were mostly less than 10%, but at  $S_{\text{eff}} \leq 0.1\%$  they exceeded 40%. Thus, careful experimental calibration is required for high-accuracy CCN measurements – especially at low  $S_{\text{eff}}$ . A comprehensive comparison and uncertainty analysis of the various Köhler models and thermodynamic parameterizations commonly used in CCN studies showed that the relative deviations between different approaches are as high as 25% for (NH<sub>4</sub>)<sub>2</sub>SO<sub>4</sub> and 12% for NaCl. The deviations were mostly caused by the different parameterizations for the activity of water in aqueous solutions of the two salts. To ensure comparability of results, we suggest that CCN studies should always report exactly which Köhler model equations and parameters were used. Provided that the Aerosol Inorganics Model (AIM) can be regarded as an accurate source of water activity data for highly

[Search ACP](#)

Library Search

Author Search

[News](#)

- [Sister Journals AMT & GMD](#)
- [Financial Support for Authors](#)
- [Journal Impact Factor](#)
- [Public Relations & Background Information](#)

[Recent Papers](#)[01 | ACPD, 03 Nov 2008: Anthropogenic influence on SOA and the resulting radiative forcing](#)[02 | ACPD, 03 Nov 2008: Evidence of mineral dust altering cloud microphysics and precipitation](#)[03 | ACPD, 03 Nov 2008: Technical Note: A new method for the Lagrangian tracking of pollution plumes from source to receptor using gridded model output](#)[04 | ACPD, 03 Nov 2008:](#)

dilute solutions of  $(\text{NH}_4)_2\text{SO}_4$  and NaCl, only Köhler models that are based on the AIM or yield similar results should be used in CCN studies involving these salts and aiming at high accuracy. Experiments with  $(\text{NH}_4)_2\text{SO}_4$  and NaCl aerosols showed that the conditions of particle generation and the shape and microstructure of NaCl particles are critical for their application in CCN activation experiments (relative deviations up to 18%).

▣ [Final Revised Paper](#) (PDF, 998 KB) ▣ [Supplement](#) (11 KB) [Discussion Paper](#) (ACPD)

Citation: Rose, D., Gunthe, S. S., Mikhailov, E., Frank, G. P., Dusek, U., Andreae, M. O., and Pöschl, U.: Calibration and measurement uncertainties of a continuous-flow cloud condensation nuclei counter (DMT-CCNC): CCN activation of ammonium sulfate and sodium chloride aerosol particles in theory and experiment, *Atmos. Chem. Phys.*, 8, 1153-1179, 2008. ▣ [Bibtex](#) ▣ [EndNote](#) ▣ [Reference Manager](#)