

论著

双源CT在主动脉夹层腔内隔绝术后的应用及辐射剂量研究

王利伟, 顾建平, 姜亮, 殷信道, 许权

南京医科大学附属南京医院医学影像科, 南京 210006

摘要:

目的: 探讨双源CT在Standford B型主动脉夹层术后随访中的价值, 并比较双能量与单能量扫描模式的辐射剂量。方法: 共40名主动脉夹层术后患者行双源CT扫描。双能量组20例, 采用100 kV和140 kV, 螺距1.0; 单能量组20例, 采用120 kV, 螺距2.0。两组均采用CareDose4D自动追踪电流法。观察与分析支架状况、有无内漏、血栓吸收程度等情况。记录双能量组及单能量组的辐射剂量、采集时间、扫描长度, 观察两种扫描模式的差异。结果: 双能量组和单能量组扫描长度差异无统计学意义($P>0.05$); 双能量组采集时间明显长于单能量组($P<0.05$); 单能量组受辐射剂量明显低于双能量组($P<0.05$)。结论: 双源CT可以客观地评价主动脉夹层腔内隔绝术后情况, 全面观察术后疗效和并发症。单能量中等螺距扫描模式辐射剂量低于双能量模式。

关键词: 主动脉夹层 血管造影术 腔内隔绝术 体层摄影术 X线计算机

Application and radiation dose research of dual-source CT in aortic dissection after endovascular exclusion

WANG Liwei, GU Jianping, JIANG Liang, YIN Xindao, XU Quan

Department of Radiology, Affiliated Nanjing Hospital, Nanjing Medical University, Nanjing 210006, China

Abstract:

Objective: To evaluate dual-source CT in the Standford B aortic dissection in the postoperative follow-up and to compare radiation dose of dual-energy mode with single-energy scanning. Methods: A total of 40 patients with aortic dissection after endovascular exclusion underwent dualsource CT. Tweny patients took the dual-energy scanning mode (100 kV and 140 kV, pitch 1.0); the other 20 patients had single-energy mode (120 kV, pitch 2.0). In both groups CareDose4D technology was used to track the current method. All images were analyzed to observe the stent shape, with or without endoleaks and the rate of thrombosis absorption. Radiation dose, scan length, and scan time were recorded for all patients, and then compared between dual-energy mode and single-energy scanning. Results: The difference in scan length and image quality was not significant between the dualenergy group and the single-energy group. The acquisition time of the dual-energy group was longer than that of the single-energy group, with significant difference. Radiation dose of thesingle-energy group was lower than that of the dual-energy group, with significant difference. Conclusion: Dual-source CT can objectively evaluate the conditions after the endovascular exclusion of aortic dissection. The operative effect and complications may be analyzed comprehensively. The radiation dose of the single-energy scan mode with medium pitch is lower than that of the dual-energy mode.

Keywords: aortic dissection angiography endovascular exclusion tomography, X-ray computed

收稿日期 2013-07-30 修回日期 网络版发布日期

DOI: 10.11817/j.issn.1672-7347.2014.01.010

基金项目:

通讯作者: 顾建平, Email: ab723y@163.com

作者简介: 王利伟, 博士, 副主任医师, 主要从事心血管疾病方面的研究。

作者Email: Email: ab723y@163.com

参考文献:

1. McMahon MA, Squirrell CA. Multidetector CT of Aortic Dissection: A Pictorial Review [J]. Radiographics, 2010, 30(2): 445-460.
2. 林婷婷, 邓克学. 64排螺旋CT血管造影技术在主动脉夹层术

扩展功能

本文信息

- Supporting info
- PDF(1126KB)
- [HTML全文]
- 参考文献[PDF]
- 参考文献

服务与反馈

- 把本文推荐给朋友
- 加入我的书架
- 加入引用管理器
- 引用本文
- Email Alert
- 文章反馈
- 浏览反馈信息

本文关键词相关文章

- 主动脉夹层
- 血管造影术
- 腔内隔绝术
- 体层摄影术
- X线计算机

本文作者相关文章

PubMed

前诊断及术后随访中的价值及优越性

[J]. 安徽医科大学学报, 2012, 47(3): 305-307.

LING Tingting, DENG Kexue. The value of 64-detector spiral CT angiography in the preoperative diagnosis and postoperative followup of aortic dissection

[J]. Acta Universitatis Medicinalis Anhui, 2012, 47(3): 305-307.

3. 唐滔, 申康军, 唐浩, 等. 新型三分支型主动脉弓腔内覆膜支架在DeBakey I型主动脉夹层中的应用

[J]. 中南大学学报: 医学版, 2012, 37(7): 706-710.

TANG Tao, SHEN Kangjun, TANG Hao, et al. A new triple-branched aortic arch covered stent graft in DeBakey Type I aortic dissection

[J]. Journal of Central South University. Medical Science, 2012, 37(7): 706-710.

4. Numburi UD, Schoenhagen P, Flamm SD, et al. Feasibility of dualenergy CT in the arterial phase: Imaging after endovascular aortic repair

[J]. AJR Am J Roentgenol, 2010, 195(2): 486-493.

5. Thrumurthy SG, Karthikesalingam A, Patterson BO, et al. The diagnosis and management of aortic dissection

[J]. BMJ, 2012, 344(11): 37-42.

6. Shiga T, Wajima Z, Apfel CC, et al. Diagnostic accuracy of transesophageal echocardiography, helical computed tomography, and magnetic resonance imaging for suspected thoracic aortic dissection: systematic review and meta-analysis

[J]. Arch Intern Med, 2006, 166(13): 1350-1356.

7. Flors L, Leiva-Salinas C, Norton PT, et al. Endoleak detection after endovascular repair of thoracic aortic aneurysm using dual-source dualenergy CT: suitable scanning protocols and potential radiation dose reduction

[J]. AJR Am J Roentgenol, 2013, 200(2): 451-460.

8. Wang DJ, Fan FD, Wang Q, et al. Preliminary characterization of acute aortic dissection in the mainland of China

[J]. Chin Med J (Engl), 2011, 124(11): 1726-1730.

9. Li Y, Yang N, Duan W, et al. Acute aortic dissection in China

[J]. Am J Cardiol, 2012, 110(7): 1056-1061.

10. Johnson TR, Nikolaou K, Becker A, et al. Dual-source CT for chest pain assessment

[J]. Eur Radiol, 2008, 18(4): 773-780.

11. Shaida N, Bowden DJ, Barrett T, et al. Acceptability of virtual unenhanced CT of the aorta as a replacement for the conventional unenhanced phase

[J]. Clin Radiol, 2012, 67(5): 461-467.

12. Apfaltrer P, Hanna EL, Schoepf UJ, et al. Radiation dose and image quality at high-pitch CT angiography of the aorta: intraindividual and interindividual comparisons with conventional CT angiography

[J]. AJR Am J Roentgenol, 2012, 199(6): 1402-1409.

13. Bolen MA, Popovic ZB, Tandon N, et al. Image quality, contrast enhancement, and radiation dose of ECG-triggered high-pitch CT versus non-ECG-triggered standard-pitch CT of the thoracoabdominal aorta

[J]. AJR Am J Roentgenol, 2012, 198(4): 931-938.

本刊中的类似文章

1. 胡文, 周新民, 刘巍, 胡建国. 72例升主动脉瘤的外科治疗[J]. 中南大学学报(医学版), 2008,33(02): 183-185

2. 李亚军; 沈树斌; 吴汉江; 罗建光; 伍玉枝; 史利华; 黄生高; 李文辉; 罗建国; 聂鑫; .牙科CT成像技术在埋伏阻生牙诊治中的应用[J]. 中南大学学报(医学版), 2002,27(4): 370-

3. 段华新; 梁昌华; 邓豪余; 李新辉; 胡硕; 戴幼艺; 杨喜春; . ~(99m)Tc-MIBI心肌灌注断层显像监测

噁环类药物的心脏毒性[J]. 中南大学学报(医学版), 2003,28(2): 167-

4. 陈伟¹, 刘进康¹, 陈琼², 李文政¹, 熊曾¹, 龙学颖¹. Bayes分析指导孤立性肺结节的临床决策[J]. 中南大学学报(医学版), 2009,34(05): 401-405

5. 肖立志, 李亚军, 胡中军, 孙剑宁, 司徒卫军. 颞骨体层术在人工耳蜗植入术后的应用[J]. 中南大学学报(医学版), 2009,34(07): 672-674

6. 曹觉¹, 杨昂¹, 龙学颖¹, 刘慧¹, 曹杰妮², 李慧¹. CT肝容积测量结合CT灌注成像评价肝功能储备的应用价值[J]. 中南大学学报(医学版), 2007,32(03): 422-426

7. 龙学颖, 曹觉, 周勉, 朱智明. 不同注射流率下胰腺、主动脉强化的群体特征及个体差异对比分析

[J]. 中南大学学报(医学版), 2007,32(06): 1064-1068

8. 龙学颖¹, 李宜雄², 王宪伟², 欧阳洋², 曹觉¹. 胰岛素瘤术前CT定位诊断的回顾性分析[J]. 中南大学学报(医学版), 2009,34(02): 165-171

9. 龙学颖, 曹觉, 石林波, 李文政, 刘慧. 肝细胞癌MSCT灌注成像及其与survivin表达、MVD及分化程度的相关性[J]. 中南大学学报(医学版), 2009,34(11): 1096-1102

10. 刘进康¹, 熊曾¹, 胡成平², 周漠玲¹, 周晖¹, 陈伟¹, 夏宇¹. 肺癌CT灌注成像与微血管管腔化程度的相关性[J]. 中南大学学报(医学版), 2010,35(12): 1242-

11. 唐朝珍;. 吡啶花青绿造影对外伤性脉络膜破裂的观察[J]. 中南大学学报(医学版), 2000,25(5): 510-

12. 江杰; 杨作成; 陈淳媛;. 儿童主动脉夹层动脉瘤2例[J]. 中南大学学报(医学版), 2001,26(4): 293-

13. 伍玉枝; 李德泰; 杨树仁; 罗建光; 谭利华; 张子曙;. 上颌窦黏液囊肿的CT诊断[J]. 中南大学学报(医学版), 2001,26(4): 356-

14. 陈建强, 韩向君, 史克珊, 关莹, 战跃福, 李小华, 李香营, 聂柳. 兴趣区高分辨率三维增强磁共振血管成像评价窦旁脑膜瘤[J]. 中南大学学报(医学版), 2012,37(2): 203-207

15. 唐滔, 申康军, 唐浩, 周新民, 杨进福. 新型三分支型主动脉弓腔内覆膜支架在 DeBakey I型主动脉夹层中的应用[J]. 中南大学学报(医学版), 2012,37(7): 706-710