

[\[Print Version\]](#)

[\[PubMed Citation\]](#) [\[Related Articles in PubMed\]](#)

*The Angle Orthodontist*: Vol. 54, No. 1, pp. 36–54.

## Imaging Characteristics of New Screen/Film Systems for Cephalometric Radiography

PANOS P. FATOUROS; S. JULIAN GIBBS;<sup>a</sup> STANLEY E. SKUBIC; GOPALA U. V. RAO

<sup>a</sup>Dr. S. Julian Gibbs, Dept. of Radiology & Radiological Sciences, Vanderbilt University, Nashville, TN 37232

### ABSTRACT

The advent of rare-earth screen/film combinations has brought substantial x-ray dose reductions with minimal loss in imaging quality. The physical characteristics of these recording systems are described and their potential for use in routine cephalometric studies is evaluated.

Dr. Fatouros is Associate Professor of Radiology, Medical College of Virginia in Richmond. He holds a Ph.D. degree in Physics from Ohio State University

Dr. Gibbs is Associate Professor of Radiology and Dentistry, School of Medicine, Vanderbilt University in Nashville, Tennessee. He holds a D.D.S. degree from Emory University and a Ph.D. in Radiation Biology from the University of Rochester

Mr. Skubic is a Graduate Student in Radiology at The Medical College of Virginia. He holds a M.A. degree in Physics from Duke University

Dr. Rao is Professor of Radiology at the Medical College of Virginia in Richmond. He holds an Sc.D. degree in Radiological Physics from Johns Hopkins University

**KEY WORDS:** cephalometrics, radiation dose, radiation risk, radiographic imaging systems, x-ray.